


检验检测报告

Test Report

[2020] WSZ FHL 第 1926 号

共 2 页 第 1 页

产品名称 Product name	Single-use Medical Face Mask (Medical Grade Face Mask)	规格型号 Specification	Plane Type , 175×95mm
		商 标 Brand	Carhon
委托单位/地址/联系电话 Applicant/Add/Tel	Jiangyin Hongmeng Rubber and Plastic Product Co.,Ltd./No. 166Hetun Road, Lingang Street, Jiangyin City Jiangsu, China. / 13961615651		
生产单位/地址/联系电话 Manufacturer/Add/Tel	Jiangyin Hongmeng Rubber and Plastic Product Co.,Ltd./No. 166Hetun Road, Lingang Street, Jiangyin City Jiangsu, China. / 13961615651		
样品等级 Sample grade	—	样品编号 Sample number	GW1926-2020
样品数量 Sample quantity	40 Pieces	样品接收日期 Receiving date of sample	2020/03/09
检验检测类别 Test type	Entrusted Inspection	货号/批号/款号 Article number/Batch number/Style number	20200301
检验检测日期 Test date	2020/03/12~2020/03/14	检验检测地点 Test site	Inspection Room
样品状态 Sample state	In Compliance With Inspection and Inspection Requirements		
检验检测依据 Test standard(s)	YY 0469-2011 <i>Surgical Mask</i> YY/T 0969-2013 <i>Single-use Medical Face Mask</i> Q/JYJH 001-2020 <i>Single-use Mask</i>		
检验检测项目 Test items	Appearance, Structure and Size, Nose Clip, Mask Belt, Ventilation Resistance, Bacterial Filtration Efficiency		
检验检测结论 Test conclusion	The samples have been inspected, all of which are in compliance with YY/T 0969-2013. Please refer to the page 2 for the specific inspection propose. Date : 2020/3/18		
备 注 Note	All of the inspection projects as required by the client are fully in accordance with YY/T 0969-2013 disposable medical mask. The samples are all provided by the client. This inspection report is served pertaining to the such samples here of.		

批准: 陈冰

审核: 李敏娟

主检: 杨莹

检验检测结果

Test Result

[2020] WSZ FHL 第 1926 号

共 2 页 第 2 页

序号 Number	检验检测项目 Test item	单位 Unit	技术要求 Technical requirement	检验检测结果 Test result	单项评价 Single Item decision		
1	Appearance*	—	The appearance of the mask should be clean, the shape is intact, the surface must not be damaged and clean.	The appearance of the mask is clean, the shape is intact, the surface is clean and is not damaged.	Qualified		
2	Structure and Size*	—	After wearing the mask, it should be able to cover the wearer's mouth, nose and jaw. The size shall conform to the designed size of 175mm×95mm, and the maximum deviation shall not exceed ±5%.	After wearing the mask, it is able to cover the wearer's mouth, nose and jaw.	Qualified		
				Sample		Actual Size/mm	Deviation/%
				Sample 1		Length:173 Width:96	Length:-1.1 Width:1.1
				Sample 2		Length:172 Width:97	Length:-1.7 Width:2.1
Sample 3	Length:173 Width:96	Length:-1.1 Width:1.1					
3	Nose Clip*	— cm	Masks should be fitted with nose clips made of plastic materials. The length of nose clip should be not less than 8.0.	Masks are fitted with nose clips made of plastic materials.	Qualified		
				9.1			
				9.2			
4	Mask Belt*	— N	Mask belt should be easily worn. The breaking force at the connection point between each mask belt and the mask body should be not less than 8.	Mask belt is easily worn.	Qualified		
				27.2			
				28.6			
5	Ventilation Resistance	Pa/cm ²	The resistance to air exchange between the two sides of the mask should be not greater than 49.	28.1	Qualified		
				10.8			
				10.6			
6	Bacterial Filtration Efficiency	—	≥95%	10.4	Qualified		
				96.8%			
				97.2%			
				97.4%			
Note	The inspection project refers to YY/T 0969-2013.						

以下空白